



Association of  
American Plant  
Food Control  
Officials

# Magruder Certificate of Analysis

Magruder Sample #220851  
Grade 0-0-50

| Analyte                        | Value | ± Uncertainty | # Labs |
|--------------------------------|-------|---------------|--------|
| Soluble K <sub>2</sub> O (50%) | 52.02 | 0.41          | 17     |

1. Data obtained from participant results in the Magruder Check Sample Program.
2. Guaranteed concentrations noted by the value preceding the unit in Analyte name.
3. Values are robust means. Uncertainty determined as  $1.25 \times \text{robust standard deviation} / \sqrt{\text{\#Labs}}$ .  
Results only shown for analytes tested by 16 or more labs. Values and uncertainties for analytes tested by less than 16 labs are available in Analyte Reports.