



# Magruder Certificate of Analysis

## Magruder Sample #221215 KCl

Analyte	Value	± Uncertainty	# Labs
Soluble K <sub>2</sub> O (62%)	61.42	0.20	82
Water Soluble Cl (%)	47.63	0.22	20
Sodium (%)	1.225	0.022	21

1. Data obtained from participant results in the Magruder Check Sample Program.
2. Guaranteed concentrations noted by the value preceding the unit in Analyte name.
3. Values are robust means. Uncertainty determined as  $1.25 \times \text{robust standard deviation} / \sqrt{\text{\#Labs}}$ .  
Results only shown for analytes tested by 16 or more labs. Values and uncertainties for analytes tested by less than 16 labs are available in Analyte Reports.