



# Magruder Certificate of Analysis

## Magruder Sample #250911

Analyte	Value	± Standard Uncertainty	# Labs
Total N (46%)	46.32	0.06	66

1. Data obtained from participant results in the Magruder Fertilizer PT Program.
2. Guaranteed concentrations noted by the value preceding the unit in Analyte name.
3. Values are robust means. Standard uncertainty determined as  $1.25 \times \text{robust standard deviation} / \sqrt{\text{\#Labs}}$ . Results only shown for analytes tested by 16 or more labs. Values and uncertainties for analytes tested by less than 16 labs are available in the Analyte Summary Statistics Report.