



Magruder Certificate of Analysis

Magruder Sample #250631
Grade 32-0-0 UAN

Analyte	Value	± Uncertainty	# Labs
Not enough labs participated for a Certificate of Analysis on any analyte.			

1. Data obtained from participant results in the Magruder Check Sample Program.
2. Guaranteed concentrations noted by the value preceding the unit in Analyte name.
3. Values are robust means. Uncertainty determined as the robust standard deviation / $\sqrt{\text{\#Labs}}$. Results only shown for analytes tested by 16 or more labs. Values and uncertainties for analytes tested by less than 16 labs are available in Analyte Reports.